

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/2/2014

DBH
mmB

Instrument Serial Number: 010440 As of 02-Jun-14

Start Date	End Date	Agency Name
30-May-14		Lynchburg PD
20-May-14	30-May-14	DFS Central Lab
22-Apr-09	20-May-14	Buena Vista PD
13-Nov-08	22-Apr-09	DFS Central Lab
05-Nov-08	13-Nov-08	Charlotte County SO
13-May-08	05-Nov-08	DFS Central Lab

DOB
mm B

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010440 Worksheet Start Date 5/30/2014
Location Lynchburg Jail
Address 907 Clay St Lynchburg VA 24504
DFS Technician Donald Hall License No. 21250

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 747 Reference Barometer (mm HG) 747

Reference Barometer(RB)Serial # 009109 RB Calibration Due 8/1/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.299
0.300	0.295	0.286	0.303	Sample 2	0.297
Precision		sample min	sample max	Sample 3	0.297
0.002		0.297	0.299		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.098	0.095	0.101	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.099
0.001		0.098	0.099		

Dry gas standard Lot No. (with tank no.)

AG406301-25

☐ Replaced dry gas standard (+O-ring)

☒ Installed at Location

☐ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010440

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Douglas B. Hall

Date

6/2/14

Issuing Analyst

Michelle B. Allen

Date

*6/4/14**msk*

Intox EC/IR-II: Accuracy Check

DBH
MMB

Department of Forensic Science

Serial Number: 010440 Test Number: 531

Test Date: 05/30/2014 Test Time: 12:01 EDT

Dry Gas Target: 0.295

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 497 psi Barometric Pressure: 747 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:02
CHK	0.299	12:02
BLK	0.000	12:04
CHK	0.297	12:05
BLK	0.000	12:06
CHK	0.297	12:07

Test Status: *Success*

Calibration CRC: 4A342050

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010440 Test Number: 532

Test Date: 05/30/2014 Test Time: 12:16 EDT

Dry Gas Target: 0.098

Lot Number: AG406301-25 Exp Date: 03/04/2016

Tank Pressure: 1120 psi Barometric Pressure: 747 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	12:16
CHK	0.098	12:17
BLK	0.000	12:19
CHK	0.098	12:20
BLK	0.000	12:21
CHK	0.099	12:22

Test Status: Success

Calibration CRC: 4A342050

JBH
mmB

DBH
hmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED	NAME OF COURT
INSTRUMENT, TEST,	DFS

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY
HALL, DONALD, B		DFS Central Lab
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED
21250	10/01/2014	05/30/2014
TEST EQUIPMENT NUMBER		
010440		

RESULTS: TIME SAMPLE TAKEN 12:31 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 02 Jun 2014 at 11:46

DBH
mmb

Test Results

Instrument Serial Number 010440

Test # 000533 Subject Test

Test Location 1 Department of
Test Date 30 May 2014

Test Location 2 Forensic Science
Test Time 12:25
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 30 May 2014 End Time 12:32

Result Time 12:31

Result Date 30 May 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:25

Data Type BLK

Sample Value 0.000

Sample Time 12:26

Data Type CHK

Sample Value 0.098

Sample Time 12:27

Data Type BLK

Sample Value 0.000

Sample Time 12:28

Data Type SUBJ

Sample Value 0.000

Sample Time 12:29

Data Type BLK

Sample Value 0.000

Sample Time 12:30

Data Type SUBJ

Sample Value 0.000

Sample Time 12:31

Data Type BLK

Sample Value 0.000

Sample Time 12:32

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG406301-25

Standard Expiration Date 03/04/2016

Tank Pressure 1120

Barometric Pressure 746 mmHg

Blow Sample Number 1 Blow Duration 4.05 sec

Blow Volume 2168 cc End-of-Blow Time 12:29

Blow Sample Number 2 Blow Duration 4.08 sec

Blow Volume 2175 cc End-of-Blow Time 12:31

Tamper Evident Stamp 8a8e3fd1

Test Status Code 0

Test Status Success